

BIOLOGICAL INDICATOR COLOR CONVERSION CHART



**AX100 - 48 hour
Formaldehyde BI**

CHEMICAL INDICATOR

PRIOR STERILIZATION

AXIS AX 100

Geobacillus stearothermophilus

Ster. | Load | Date

Lot

AFTER STERILIZATION

AXIS AX 100

Geobacillus stearothermophilus

Ster. | Load | Date

Lot

CULTURE MEDIUM

Negative(-)



Positive(+)



THE COLOR ON THE PRODUCT CAN SHOW DIFFERENCE WITH THE COLOR SEEN ON REFERENCE CHART DUE TO PRINTING QUALITY. USER CAN MAKE COMPARISON WITH UNUSED PRODUCT.